



Ingeniería y Servicios Metrología Tridimensional S.L.  
33203 Gijón Asturias  
España  
VAT No ESB33951526  
Tel.: +34 984 194 467  
[www.ism3d.es](http://www.ism3d.es)  
[info@ism3d.es](mailto:info@ism3d.es)

**OBJETO** : **Dumbbell with random black and white pattern**  
*Item*

**MARCA** : **Trapet PE**  
*Mark*

**MODELO** : **Dumbbell L125 D75**  
*Model*

**IDENTIFICACIÓN** : **10-2019**  
*Identification*

**SOLICITANTE** : **PHOTOGAUGE**  
*Applicant* **Chennai 600042**  
**India**

**FECHA/S DE CALIBRACIÓN** : **17.3.2019**  
*Date/s of calibration*

**Signatario/s autorizado/s**  
*Authorized signatory/ies*



**Dr.-Ing. Eugen Trapet**  
Jefe de Laboratorio

**Fecha de emisión**  
*Date of issue*  
**17.3.2019**

Este certificado se expide de acuerdo con las condiciones de la acreditación concedida por ENAC, que ha comprobado las capacidades de medida del laboratorio y su trazabilidad a patrones nacionales o internacionales. Este certificado no podrá ser reproducido parcialmente sin la aprobación por escrito del laboratorio que lo emite y de ENAC.

ENAC es firmante del Acuerdo de Reconocimiento Mutuo (MLA) de calibración de European Cooperation for Accreditation (EA) y de International Laboratory Accreditation Cooperation (ILAC).

This certificate is issued in accordance with the conditions of accreditation granted by ENAC which has assessed the measurements capability of the laboratory and its traceability to national or international standards. This certificate may not be partially reproduced, except with the prior written permission of the issuing laboratory and ENAC.

ENAC is one of the signatories of the Multilateral Agreement of the European Cooperation for Accreditation (EA) and the International Laboratory Accreditation Cooperation (ILAC).

## **Method**

The dumbbell (2-ball ball bar) was measured according to the procedure T 03-01 of ISM3D using a coordinate measuring machine (CMM).

The dumbbell was measured by probing each ball with 29 equally distributed points on a hemisphere. The dumbbell was measured in 4 different orientations in the measuring volume of the CMM, with the dumbbell between each orientation rotated 90° around its axis so that the probing points lay in different but overlapping areas/hemispheres; these independent measurements were corrected for temperature and averaged in case of the measurands “length” and “diameter”. In case of the measurand “form error”, the worst (biggest) range of radial deviations is reported as result.

Linear scale factor and probe tip diameter were not corrected, as the above specified procedure does not require a correction.

## **Coordinate system und alignment**

No coordinate system is required to state the results.

Ball 1 is the one closest to the dot mark on the stem.

## **Environmental conditions**

The reference temperature is 20°C. All measured lengths are referred to 20°C. The temperature of the environment was kept at 20.3°C +/- 0.3K during the calibration. For the linear thermal expansion coefficient of the dumbbell length a value of 0.0000152 has been used, for the sphere diameters a value of 0.000022 1/K. The value is based on calculation from nominal values: steel shaft in aluminium spheres, 20 mm deep.

## Uncertainty

The uncertainty  $U$  (for  $k=2$ ) of the center to center distance is  $U = 1.70 \mu\text{m}$  and of the diameters  $U = 2.00 \mu\text{m}$ . The uncertainty of the deviations from the substitute (ideal form) Gaussian sphere is  $1.80 \mu\text{m}$  (note: the uncertainty of the range of deviations is thus  $2.55 \mu\text{m}$ ).

This uncertainty of length and diameter are limited due to the form error of the object. They are bigger than the accredited scope of ISM3D. The analysis of the uncertainty of the deviations of form has been performed according to the Method of the Virtual CMM as described in procedure T3-01 "Calibración de piezas patrón con MMC Virtual" and T1-06 "Calibración de los parámetros de la MMC Virtual".

The expanded uncertainty of the measurement results has been obtained by multiplying the standard uncertainty by the coverage factor  $k>2$ , due to the small number of measurements, which for a standard distribution of values results approximately in a probability of coverage of 95%. The standard uncertainty has been determined in accordance with A-4/02 M: 2013 and ILAC-P14.

The uncertainties associated with the results correspond to the moment of the calibration, they don't contain provisions for long term variations of the object. They are only valid for the object identified on the first page of this certificate.

## Results

Center distance of the balls: **123.84185 mm**  
Diameter of ball 1 (near the white dot): **75.58414 mm**  
Diameter of ball 2: **75.72714 mm**  
Range of radial deviations from Gaussian Substitute Sphere of ball 1: **0.00996 mm**  
Range of radial deviations from Gaussian Substitute Sphere of ball 2: **0.00971 mm**